



Patient : TOPCON DEMO

Patient ID :

Date Of Birth : 01/01/1950  
(mm/dd/yyyy)

**OD**

Phakic

Data Measurements n: 1.3375

Aladdin Optical

AL : 23.73 mm K1 : 8.28 mm @ 8°  
 ACD : 3.14 mm K2 : 8.00 mm @ 98°  
 LT 4.04 mm CYL : -1.45 D ax 8°  
 CCT 0.544 mm

Target Refraction: 0

Oculentis L-313

SRK/T	
IOL(D)	REF(D)
20.50	0.83
21.00	0.47
<b>21.50</b>	<b>0.10</b>
22.00	-0.27
22.50	-0.64

IOL @ Target A = 118.100  
21.64

Oculentis LS-313 MF30

SRK II	
IOL(D)	REF(D)
21.00	0.77
21.50	0.37
<b>22.00</b>	<b>-0.03</b>
22.50	-0.43
23.00	-0.83

IOL @ Target A = 118.600  
21.97

Oculentis LU-313 MF30T

Haigis	
IOL(D)	REF(D)
21.50	0.58
22.00	0.21
<b>22.50</b>	<b>-0.16</b>
23.00	-0.54
23.50	-0.92

IOL @ Target A0 = 0.870  
22.28 A1 = 0.400  
A2 = 0.100

Oculentis LS-412Y

Hoffer Q	
IOL(D)	REF(D)
21.00	0.86
21.50	0.51
<b>22.00</b>	<b>0.16</b>
22.50	-0.20
23.00	-0.56

IOL @ Target pACD = 5.070  
22.22

Oculentis LU-800 RZI

Holladay I	
IOL(D)	REF(D)
19.00	0.90
19.50	0.52
<b>20.00</b>	<b>0.13</b>
20.50	-0.25
21.00	-0.65

IOL @ Target SF = 0.310  
20.17

Topcon Aladdin Biometer

Surgeon : SURGEON GENERIC

Exam Date : 02/10/2015 - 17:55  
(mm/dd/yyyy)

**OS**

Phakic

Data Measurements n: 1.3375

Aladdin Optical

AL : 23.93 mm K1 : 8.51 mm @ 173°  
 ACD : 3.21 mm K2 : 7.90 mm @ 83°  
 LT 4.00 mm CYL : -3.06 D ax 173°  
 CCT 0.556 mm

Target Refraction: 0

Oculentis L-313

SRK/T	
IOL(D)	REF(D)
20.50	0.67
21.00	0.31
<b>21.50</b>	<b>-0.06</b>
22.00	-0.43
22.50	-0.81

IOL @ Target A = 118.100  
21.42

Oculentis LS-313 MF30

SRK II	
IOL(D)	REF(D)
21.00	0.62
21.50	0.22
<b>22.00</b>	<b>-0.18</b>
22.50	-0.58
23.00	-0.98

IOL @ Target A = 118.600  
21.77

Oculentis LU-313 MF30T

Haigis	
IOL(D)	REF(D)
21.00	0.81
21.50	0.45
<b>22.00</b>	<b>0.08</b>
22.50	-0.30
23.00	-0.67

IOL @ Target A0 = 0.870  
22.10 A1 = 0.400  
A2 = 0.100

Oculentis LS-412Y

Hoffer Q	
IOL(D)	REF(D)
21.00	0.72
21.50	0.37
<b>22.00</b>	<b>0.01</b>
22.50	-0.35
23.00	-0.71

IOL @ Target pACD = 5.070  
22.02

Oculentis LU-800 RZI

Holladay I	
IOL(D)	REF(D)
19.00	0.76
19.50	0.38
<b>20.00</b>	<b>-0.01</b>
20.50	-0.40
21.00	-0.80

IOL @ Target SF = 0.310  
19.99